

Dr. K. S. Lokesh
Director, CEMARF
Mobile : 9035500208
Office : 08392-242915
Fax : 08392-242806
E- mail : lokeshsk@gmail.com



Vijayanagara Sri Krishnadevaraya
University
CEMARF
JnanaSagara Campus
Vinayakanagara, Near Cantonment
BALLARI - 583 105, KARNATAKA

No: VSKUB/CEMARF/2023-24/92

Date: 15 /03/2024

*****Notification*****

(Through Proper Channel)

Sub: Request to submit the quotation.


Ref: No. VSKUB/ADM/Gen/2023-24/8350, dated: 14/03/2024

With respect to the above subject, we need Micro Weighing balance for CEMARF. In this regard, I request you to send the quotation for Micro Weighing balance before 21/03/2024. The specification of Micro Weighing Balance is attached for your kind perusal and needful action.

Sl.No	Equipment	Qty (No)	Specification
1	Micro Weighing balance		Micro Weighing balance: Range: 0.1mg to 10g, Linearity: 0.2mg, Response time: 3s, Readability: =0.1mg, Repeatability: ≤0.1mg. Calibration: internal, user programmable units, Application Modes: Mass unit conversion by toggling, tare memory, net total, weighing in percent, counting, averaging Construction: Draft Shield Chamber from pan to top of glass door

Thanking you

Sincerely Yours


(Dr.K.S.Lokesh)

Dr. K.S. LOKESH M.Sc.,Ph.D.,
Director & Professor, CEMARF
Vijayanagara Sri Krishnadevaraya University
BALLARI-583 105. Karnataka.

Copy to:

- University Notice board, VSKU Ballari.
- Director, ICT, VSK University, Ballari with a request for uploading the same in university website.
- Office copy, VSK University, Ballari.